

**Central Stores**  
**Directorate Medical Delhi**  
**ESI Dispensary Complex**  
**Tilak Vihar , New Delhi-110018**

DMU-16/55/CPAP/2008/CS

Dated:08.11.2011

To

\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

Sub:Limited tender enquiry for procurement of Two Nasal CPAP and One Auto CPAP.

Sir,

Director(Medical)Delhi ,ESI Dispensary complex Tilak Vihar intends to purchase Two Nasal CPAP and One Auto CPAP from authorized dealers/distributor /manufacturer.

You are therefore ,requested to submit the rate enquiry in the enclosed Annexure "A"(Technical Bid) and Annexure "B" (Price bid).The technical bid and price bid should be submitted in separate sealed envelope .Open price bids will not be considered. The last date of submission of tender enquiry is 25.11.2011 by 1:00PM to Central Stores ESI Dispensary complex Tilak Vihar,New Delhi-110018 superscribing as **"RATE ENQUIRY FOR Two Nasal CPAP and One Auto CPAP .**

Terms & conditions:-

- 1)EMD of Rs 5000.00 in the form of Demand Draft/ Banker's Cheque in favour of ESIC Fund Acc No. 1 should be submitted along with tender document. EMD of unsuccessful tenderer will be returned without any interest after finalization of tender process. EMD of successful tender will be returned without interest after receiving the goods and inspection of said item.
- 2)Copy of VAT/Sales tax registration certificate.
- 3)Rates ,sales tax must be quoted separately .
- 4)Lowest Govt./Institutional rates.
- 5)List of Institutions where supply made if any.
- 6)Warranty period should be clearly outlined .
- 7)Written assurance regarding after sales service.
- 8)AMC conditions should be clearly outlined.
- 9)Payment will be made within seven days of receipt of goods after inspection.
- 10)Delivery to be made from ex-stock within Five days of issue of supply order.
- 11)Brochure/Literature of the models quoted should be enclosed.
- 12)Tenderer will have to demonstrate the quoted model before the technical committee when asked for the same.
- 13)The procurement of number of CPAP'S/Auto CPAP may vary depending upon future demand.

Yours sincerely

Manager Stores

Encl: ANNEXURE "A" & "B"

ANNEXURE "A" (TECHNICAL BID)

ITEM NAME	Qty	SPECIFICATIONS	MODEL & BRAND
Nasal CPAP	02	1) Pressure Range : 04-20CM of H2O 2) Ramp: 0-20minutes (5 minutes increment) 3) Noise Level: Low Noise Level (Preferably <30db) 4) Should have humidifier 5) Provided with soft silicone Nasal Mask and tubings 6) Light weight 7) Automatic ON/OFF 8) Mask off Alert 9) Should run on electricity at 220V, 50-60Hz 10) Should have warranty of One year or more	
Auto CPAP	01	1) Pressure Range : 04-20CM of H2O 2) Ramp: 0-20minutes (5 minutes increment) 3) Noise Level: Low Noise Level (Preferably <30db) 4) Should have humidifier 5) Provided with soft silicone Nasal Mask and tubings 6) Light weight 7) Automatic ON/OFF 8) Mask off Alert 9) Should run on electricity at 220V, 50-60Hz 10) Should have warranty of One year or more 11) Auto adjustment pressures	

ANNEXURE "B" (PRICE BID)

ITEM	MODEL	RATE/UNIT
Nasal CPAP		
Auto CPAP		