

**Amendment No. 1 Dated 03.02.2011**  
**Ref: Tender No. HLL/PCD/ESIC-44/10-11**

**The following amendments are issued in respect of list of requirements:-**

**Tender No.: HLL/PCD/ESIC-44/10-11**

**In the list of requirement 1 item of Rohini added:**

| <b>S.NO.</b> | <b>ITEM DESCRIPTION</b>                           | <b>TOTAL QNTY.</b> | <b>EMD AMOUNT Rs.</b> |
|--------------|---|--------------------|-----------------------|
| <b>1</b>     | <b>Air condition Mortuary Cabinet (04 bodies)</b> | <b>1(Rohini)</b>   | <b>10000</b>          |

**TECHNICAL SPECIFICATION**

**Air Condition Mortuary Chamber Fully Automatic (04) Bodies**

- **Capacity : Four Bodies in two rows**
- **Temperature Range : -2 degree to 5 degree C**
- **Controller : Microprocessor Based**
- **Interior Panel : Made of Stainless Steel**
- **Outer Panel : Made of Stainless Steel**
- **Door : Hinged Insulated Doors with magnetic gaskets**
- **Insulation : Polyurethane foam with thickness of 80 – 100 mm**
- **Body Trays : Stainless Steel with Telescopic Mechanism**
- **Compressor : Heavy Duty low noise & minimal vibration**
- **Condenser : Efficient with Auto – condensing evaporating system**
- **Air Circulation: Forced Air Circulation**
- **Alarms : Audio – Visual Alarms to warn high or low temperature**
- **Supplies : 220 V/ 50Hz**
- **Illumination : Door operated illumination system in the chamber**
- **Accessories : Temperature recorder, data logger, voltage safety System, with battery backup.**